

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

JUL 07 2009

DataMaster cdm S/N 130296

Supervisor/Operator Performing the Verification Procedure:

Name John J. Waldron ID# 3392 Date 06/30/09 ✓

A.

Agency Unalaska Department of Public Safety Phone 907-581-1233

Instrument Location Unalaska Department of Public Safety, PO Box 370, 29 Safety Way, Unalaska, AK 99685 ✓

Alco S/N X172985 ✓ Target Value .083 ✓ High Pressure 500

B.

Alco Test Values .084 ✓ .085 ✓
1st Test Value 2nd Test Value

Signature [Signature] 7/14/09

(OVER)

(Do not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 20th day of Jul, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

Department and Date: Unalaska DPS,

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

[illegible]